



Declaration of Conformity
MDD 93/42/EEC
Risk Class: IIa



Application of Standard(s): IEC 601-1/ EN60601-1, Medical Electrical Standard + when applicable
 IEC 601-1-2/EN60601-1-2 Electromagnetic Compatibility, + when applicable amendments and applicable part two standards effective at the time of evaluation.
 IEC 60601-2-5 Particular Requirements for the Safety of Ultrasonic Physical Therapy Equipment
 IEC 60601-2-10 Particular Requirements for the Safety of Nerve and Muscle Stimulators

Manufacturer's Name Chattanooga Group
Manufacturer's Address A Division of Encore Medical
 4717 Adams Road
 Hixson, Tennessee 37343 U.S.A.
 Telephone: 1 + 423-870-7200

Authorized Representative's Name Medical Device Safety Service
Authorized Representative's Address Burckhardtstr. 1
 D-30163 Hannover Germany
 Telephone: +49-5103-939430

Model Name	Type of Equipment	UMDNS Code
OptiFlex®, OptiFlex ³ ®	Exercisers, Continuous Passive Motion, Lower Limb	17-138
Para-Care®	Baths, Paraffin, Physical Therapy	12-956
Intelect® Intelect ® Mobile Ultrasound- 2776, 27467	Ultrasound Units, Physical Therapy	11-248
Intelect® Intelect® Advanced Therapy System- 2762CC, 2772MC, 2762AS, 2772AS, 79976, 79977, 79966, 79967, 79968, 79969, 79970, 79971, 27468, 27465, 10648, 2775, 27312, 27313, 27314, 27315 Intelect® User Remote Control- 27508, 27079 Intelect® Patient Interrupt Switch- 27469, 27470 Intelect® Mobile Combo- 2778, 27467	Ultrasound Units/Neuromuscular Stimulators, Physical Therapy	17-908
Intelect® Intelect ® Advanced Therapy System- 2765CS, 2773MS, 2765AS, 2773AS, 79976, 79977, 79966, 79967, 79968, 79969, 79970, 79971, 27468, 27465, 10648, 2775, 27312, 27313, 27314, 27315 Intelect® Channel 3/4 Electrotherapy Module- 2770 Intelect® Mobile Stim- 2777, 27467	Stimulators, Neuromuscular, Therapeutic	16-255
Intelect®	Stimulators, Electroanalgesic, Transcutaneous Electrical Nerve	13-782
Intelect® Ultrasound Applicators- 27333, 27334, 27335, 27336	Transducers, Ultrasonic	14-121
Intelect® Battery Module- 2767	Batteries, Medical Device	16-337
Intelect® Mobile Battery Pack- 27478		




Declaration of Conformity
MDD 93/42/EEC
Risk Class: IIa



Model Name	Type of Equipment	UMDNS Code
Forte CPS	Ultrasound Units, Physical Therapy	11-248
Forte CPS	Ultrasound Units/Neuromuscular Stimulators, Physical Therapy	17-908
Forte CPS	Stimulators, Neuromuscular, Therapeutic	16-255
VitalStim®- 5900, 5905	Stimulators, Neuromuscular, Therapeutic	16-255
VitalStim® Electrodes- 59000, 59005, 59035	Electrodes, Neuromuscular Stimulator	11-454
Dura-Stick®- 42061, 42041, 42042, 42043, 42044, 42045	Electrodes, Neuromuscular Stimulator	11-454
Dura-Stick® TENS	Electrodes, Transcutaneous Electrical Nerve Stimulation	17-191
EMG Retrainer® & EMG Retrainer® IR sEMG Module- 2771, 27321, 27322	Biofeedback System	10-396
EMG Electrodes- 42109, 42140	Electrodes, Electromyographic	11-441
EMG Vaginal Probe- 77725	Probes, Other	15-237
EMG Rectal Probe	Probes, Other	15-237
Tru-Trac®, Triton®, TX- 4749, 4798, 4759, 4779, 4766	Traction Units	14-105
Tru-Trac®	Belts, Traction	14-100
Fluidotherapy® DHT	Hot Dry Air/Fluidized Medium Units	16-982
Theratherm®	Heating Pads, Electric	11-989
PresSsion®	Compression Units	16-836
Vacuum System Intelect® Vacuum Module- 2785, 2774, 56852, 56806, 27745, 27746, 27657, 27658, 56851, 56807	Vacuum Stabilized Pads	15-325
Chattanooga Ionto and Ionto Ultra-1331, 1332, 1333, 1335, 1326, 1327, 1328, 1329	Iontophoresis Units, Drug Delivery	18-155

We declare that our product models listed above meet the applicable provisions of the European Medical Devices Directive 93/42/EEC.


 DATE: 01/18/06
Michael Treas
Manager of Regulatory Affairs
Chattanooga Group,
A Division of Encore Medical